

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
 Data File : VX004250.D
 Acq On : 24 Aug 2018 16:42
 Operator : JC/MD
 Sample : J4270-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BU-03-082318

Quant Time: Aug 25 01:41:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	298239	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	477111	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	438447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	261292	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	204008	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
35) Dibromofluoromethane	5.50	113	175032	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
50) Toluene-d8	8.71	98	674464	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.14	95	242005	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.78	43	6360	1.103	ug/l	97
20) Methylene Chloride	2.85	84	76096	20.336	ug/l	98
30) Chloroform	5.21	83	6515	1.029	ug/l	94
43) Isopropyl Acetate	6.46	43	171156	21.115	ug/l	99
95) Naphthalene	13.83	128	24969	1.465	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
Data File : VX004250.D
Acq On : 24 Aug 2018 16:42
Operator : JC/MD
Sample : J4270-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BU-03-082318

Quant Time: Aug 25 01:41:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

