

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
 Data File : VX004254.D
 Acq On : 24 Aug 2018 18:28
 Operator : JC/MD
 Sample : J4621-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-1

Quant Time: Aug 25 02:06:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	310945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	501643	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	440807	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	264824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	214440	52.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.82%	
35) Dibromofluoromethane	5.50	113	185373	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	8.71	98	701568	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	11.14	95	254384	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
Target Compounds						
18) Methyl Acetate	2.78	43	8268	1.375	ug/l	98
20) Methylene Chloride	2.86	84	81012	20.765	ug/l	96
30) Chloroform	5.21	83	8502	1.288	ug/l	97
43) Isopropyl Acetate	6.46	43	181964	21.351	ug/l	100
95) Naphthalene	13.83	128	123820	7.166	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX082418\
Data File : VX004254.D
Acq On : 24 Aug 2018 18:28
Operator : JC/MD
Sample : J4621-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-1

Quant Time: Aug 25 02:06:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

