

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082419\
 Data File : VX011828.D
 Acq On : 24 Aug 2019 18:59
 Operator : JC/SP
 Sample : K4474-28
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T1-2-F4-(2.5)

Quant Time: Aug 26 04:42:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	449656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	612434	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	572226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	313799	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	201688	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
35) Dibromofluoromethane	5.49	113	180609	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
50) Toluene-d8	8.71	98	709000	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
62) 4-Bromofluorobenzene	11.13	95	258055	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
Target Compounds						
16) Acetone	2.43	43	10954	5.575	ug/l	93
20) Methylene Chloride	2.85	84	13694	3.409	ug/l #	95
43) Isopropyl Acetate	6.43	43	64182	7.802	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082419\
Data File : VX011828.D
Acq On : 24 Aug 2019 18:59
Operator : JC/SP
Sample : K4474-28
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
T1-2-F4-(2.5)

Quant Time: Aug 26 04:42:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration

