

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082419\
 Data File : VX011832.D
 Acq On : 24 Aug 2019 20:33
 Operator : JC/SP
 Sample : K4474-33
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T1-2-K2-(2.5)

Quant Time: Aug 26 04:49:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 01:48:12 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	452327	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	613916	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	581008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	316228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	203766	44.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.68%	
35) Dibromofluoromethane	5.49	113	184077	46.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.74%	
50) Toluene-d8	8.71	98	710219	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	11.13	95	262683	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
Target Compounds						
16) Acetone	2.43	43	10900	5.514	ug/l	95
20) Methylene Chloride	2.85	84	14933	3.695	ug/l	98
43) Isopropyl Acetate	6.44	43	65720	7.970	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082419\
Data File : VX011832.D
Acq On : 24 Aug 2019 20:33
Operator : JC/SP
Sample : K4474-33
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T1-2-K2-(2.5)

Quant Time: Aug 26 04:49:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 01:48:12 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. 0.01 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33

Instrument :

MSVOA_X

ClientSampleId :


T1-2-K2-(2.5)

Tot Ion:168 Resp: 452327

Ion Ratio Lower Upper




168 100

99 41.1 39.4 59.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 5.514 ug/l

RT: 2.43 min Scan# 210

Delta R.T. -0.00 min

Lab File: VX011832.D


Acq: 24 Aug 2019 20:33

Tgt Ion: 43 Resp: 10900

Ion Ratio Lower Upper



43 100


58 33.3 24.4 36.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#20
Methvlene Chloride
Concen: 3.695 ug/l
RT: 2.85 min Scan# 279
Delta R.T. 0.01 min
Lab File: VX011832.D
Acq: 24 Aug 2019 20:33




Instrument :
MSVOA_X
ClientSampleId :
T1-2-K2-(2.5)


Tot Ion: 84 Resp: 14933
Ion Ratio Lower Upper
84 100
49 108.2 88.3 132.5
51 34.3 26.2 39.2
86 62.2 50.6 76.0



#33
1,2-Dichloroethane-d4
Concen: 44.844 ug/l
RT: 6.06 min Scan# 804
Delta R.T. -0.00 min
Lab File: VX011832.D
Acq: 24 Aug 2019 20:33

Tgt Ion: 65 Resp: 203766
Ion Ratio Lower Upper
65 100
67 53.6 0.0 107.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.00 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33

Instrument :

MSVOA_X

ClientSampleId :

T1-2-K2-(2.5)


Tot Ion:114 Resp: 613916

Ion Ratio Lower Upper

114 100

63 17.2 0.0 36.8




88 14.6 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 46.369 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.00 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33


Tgt Ion:113 Resp: 184077

Ion Ratio Lower Upper

113 100

111 102.8 81.0 121.4



192 26.9 19.9 29.9




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 7.970 ug/l

RT: 6.44 min Scan# 867

Delta R.T. -0.00 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33

Instrument :

MSVOA_X

ClientSampleId :

T1-2-K2-(2.5)





Tot Ion: 43 Resp: 65720

Ion Ratio Lower Upper

43 100

61 21.8 16.2 24.2

87 14.6 11.6 17.4



#50

Toluene-d8

Concen: 51.291 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX011832.D




Acq: 24 Aug 2019 20:33


Tgt Ion: 98 Resp: 710219

Ion Ratio Lower Upper

98 100

100 65.4 52.1 78.1





#62

4-Bromofluorobenzene

Concen: 47.995 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33

Instrument :

MSVOA_X

ClientSampleId :

T1-2-K2-(2.5)


Tot Ion: 95 Resp: 262683

Ion Ratio Lower Upper

95 100

174 104.2 0.0 197.4

176 101.7 0.0 194.4




Abundance Ion 94.90 (94.60 to 95.60): VX011751.D (-1630) (-)

Ion 173.80 (173.50 to 174.50): VX011751.D (-1630) (-)

Ion 175.80 (175.50 to 176.50): VX011751.D (-1630) (-)

Time-->




Abundance Ion 94.90 (94.60 to 95.60): VX011832.D (-1588) (-)

Ion 173.80 (173.50 to 174.50): VX011832.D (-1588) (-)

Ion 175.80 (175.50 to 176.50): VX011832.D (-1588) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. -0.00 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33


Tgt Ion: 117 Resp: 581008

Ion Ratio Lower Upper

117 100

82 50.8 41.7 62.5

119 32.8 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): VX011751.D (-1462) (-)

Ion 82.00 (81.70 to 82.70): VX011751.D (-1462) (-)

Ion 118.90 (118.60 to 119.60): VX011751.D (-1462) (-)

Time-->






Abundance Ion 116.90 (116.60 to 117.60): VX011832.D (-1420) (-)

Ion 82.00 (81.70 to 82.70): VX011832.D (-1420) (-)

Ion 118.90 (118.60 to 119.60): VX011832.D (-1420) (-)

Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1791

Delta R.T. -0.00 min

Lab File: VX011832.D

Acq: 24 Aug 2019 20:33

Instrument :

MSVOA_X

ClientSampleId :

T1-2-K2-(2.5)

Tot Ion:152 Resp: 316228

Ion Ratio Lower Upper

152 100

115 52.4 37.3 111.8

150 157.4 0.0 343.0

Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

