

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082421\
 Data File : VX023921.D
 Acq On : 24 Aug 2021 11:01
 Operator : JC/MD
 Sample : VX0824WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0824WBL01

Quant Time: Aug 25 03:54:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	201401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	332200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	297690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128143	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	151628	53.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.880%
35) Dibromofluoromethane	5.397	113	112066	50.968	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.653	98	406370	48.559	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.085	95	141006	45.771	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082421\
Data File : VX023921.D
Acq On : 24 Aug 2021 11:01
Operator : JC/MD
Sample : VX0824WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0824WBL01

Quant Time: Aug 25 03:54:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

