

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037206.D
 Acq On : 24 Aug 2023 11:04
 Operator : JC/MD
 Sample : 04047-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20230816

Quant Time: Aug 24 23:37:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	153395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	251639	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	103871	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95943	43.384	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.760%
35) Dibromofluoromethane	5.385	113	77071	46.246	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.647	98	293737	48.590	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	111744	47.037	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037206.D
Acq On : 24 Aug 2023 11:04
Operator : JC/MD
Sample : 04047-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20230816

Quant Time: Aug 24 23:37:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

