

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037209.D
 Acq On : 24 Aug 2023 12:17
 Operator : JC/MD
 Sample : 04093-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-20230818

Quant Time: Aug 24 23:37:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	150126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248222	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95031	43.908	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.820%
35) Dibromofluoromethane	5.385	113	75872	46.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.300%
50) Toluene-d8	8.647	98	290951	48.792	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	112249	47.900	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.800%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1837	2.170	ug/l	98
20) Methylene Chloride	2.794	84	1009	0.533	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037209.D
Acq On : 24 Aug 2023 12:17
Operator : JC/MD
Sample : 04093-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-20230818

Quant Time: Aug 24 23:37:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

