

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037211.D
 Acq On : 24 Aug 2023 13:04
 Operator : JC/MD
 Sample : 04051-08DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20230816DL

Quant Time: Aug 25 00:15:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91946	44.035	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.080%
35) Dibromofluoromethane	5.385	113	73740	46.695	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.380%
50) Toluene-d8	8.647	98	278407	48.602	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.079	95	109421	48.606	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.220%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	714	0.466	ug/l #	80
12) 1,1-Dichloroethene	2.319	96	349	0.234	ug/l	85
27) cis-1,2-Dichloroethene	4.495	96	1667	0.834	ug/l	77
44) Trichloroethene	7.129	130	61298	33.595	ug/l	96
64) Tetrachloroethene	9.275	164	1182	0.746	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037211.D
Acq On : 24 Aug 2023 13:04
Operator : JC/MD
Sample : 04051-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20230816DL

Quant Time: Aug 25 00:15:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

