

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037225.D
 Acq On : 24 Aug 2023 18:43
 Operator : JC/MD
 Sample : 04051-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW181S-20230816

Quant Time: Aug 24 23:40:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	140249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232188	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91897	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88637	43.838	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.680%
35) Dibromofluoromethane	5.385	113	71482	46.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.980%
50) Toluene-d8	8.647	98	272258	48.810	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	101436	46.275	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.540%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	562	0.389	ug/l #	67
27) cis-1,2-Dichloroethene	4.489	96	1186	0.613	ug/l	83
30) Chloroform	5.086	83	1442	0.447	ug/l #	71
44) Trichloroethene	7.129	130	1446	0.814	ug/l	82
64) Tetrachloroethene	9.275	164	3469	2.289	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037225.D
Acq On : 24 Aug 2023 18:43
Operator : JC/MD
Sample : 04051-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW181S-20230816

Quant Time: Aug 24 23:40:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

