

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037226.D
 Acq On : 24 Aug 2023 19:07
 Operator : JC/MD
 Sample : 04051-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20230816

Quant Time: Aug 24 23:40:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	102821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94085	44.137	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	88.280%		
35) Dibromofluoromethane	5.385	113	75271	46.738	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.480%		
50) Toluene-d8	8.647	98	286905	49.112	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.220%		
62) 4-Bromofluorobenzene	11.079	95	109918	47.878	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.760%		
Target Compounds						
44) Trichloroethene	7.129	130	14073	7.563	ug/l	97
52) Toluene	8.726	92	1623	0.374	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037226.D
Acq On : 24 Aug 2023 19:07
Operator : JC/MD
Sample : 04051-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20230816

Quant Time: Aug 24 23:40:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

