

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037236.D
 Acq On : 25 Aug 2023 00:23
 Operator : JC/MD
 Sample : 04051-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D3-20230816

Quant Time: Aug 25 06:29:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	254540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104145	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96385	42.940	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	85.880%		
35) Dibromofluoromethane	5.385	113	77923	46.224	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.440%		
50) Toluene-d8	8.647	98	298135	48.756	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.520%		
62) 4-Bromofluorobenzene	11.079	95	112800	46.940	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.880%		
Target Compounds						
44) Trichloroethene	7.123	130	39219	20.136	ug/l	95
64) Tetrachloroethene	9.275	164	1210	0.727	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037236.D
Acq On : 25 Aug 2023 00:23
Operator : JC/MD
Sample : 04051-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D3-20230816

Quant Time: Aug 25 06:29:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

