

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037241.D
 Acq On : 25 Aug 2023 02:23
 Operator : JC/MD
 Sample : 04077-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D3-20230817

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/25/2023
 Supervised By :Semsettin Yesilyurt 08/25/2023

Quant Time: Aug 25 06:31:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	253481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	101139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97864	44.079	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	88.160%		
35) Dibromofluoromethane	5.385	113	77645	46.252	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.500%		
50) Toluene-d8	8.647	98	295725	48.564	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.120%		
62) 4-Bromofluorobenzene	11.079	95	111143	46.444	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.880%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	14784	9.082	ug/l	98
12) 1,1-Dichloroethene	2.319	96	796	0.502	ug/l	89
27) cis-1,2-Dichloroethene	4.489	96	8858	4.169	ug/l	79
44) Trichloroethene	7.129	130	31836	16.414	ug/l	96
45) 1,2-Dichloropropane	7.428	63	6678	3.959	ug/l	100
64) Tetrachloroethene	9.275	164	56615	34.209	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	1450m	0.673	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037241.D
Acq On : 25 Aug 2023 02:23
Operator : JC/MD
Sample : 04077-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D3-20230817

Quant Time: Aug 25 06:31:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/25/2023
Supervised By :Semsettin Yesilyurt 08/25/2023

