

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037246.D
 Acq On : 25 Aug 2023 04:23
 Operator : JC/MD
 Sample : 04096-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-RW11-MW01S-GW-20230816

Quant Time: Aug 25 07:45:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250081	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97195	44.280	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.560%
35) Dibromofluoromethane	5.385	113	76891	46.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	292930	48.759	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	109250	46.273	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.540%
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	2235	0.376	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037246.D
Acq On : 25 Aug 2023 04:23
Operator : JC/MD
Sample : 04096-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-RW11-MW01S-GW-20230816

Quant Time: Aug 25 07:45:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

