

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037250.D
 Acq On : 25 Aug 2023 05:59
 Operator : JC/MD
 Sample : 04095-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20230818

Quant Time: Aug 25 07:46:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	257619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98851	43.924	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.840%
35) Dibromofluoromethane	5.385	113	78886	46.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	8.647	98	302024	48.801	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.600%
62) 4-Bromofluorobenzene	11.079	95	113019	46.469	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%
Target Compounds						
30) Chloroform	5.093	83	4283	1.192	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037250.D
Acq On : 25 Aug 2023 05:59
Operator : JC/MD
Sample : 04095-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20230818

Quant Time: Aug 25 07:46:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

