

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037252.D
 Acq On : 25 Aug 2023 06:48
 Operator : JC/MD
 Sample : 04095-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-20230818

Quant Time: Aug 25 07:47:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258782	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	108200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97540	43.268	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.540%
35) Dibromofluoromethane	5.385	113	79289	46.264	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.520%
50) Toluene-d8	8.647	98	302515	48.661	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.079	95	114212	46.749	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	681	0.375	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.331	101	5009	3.030	ug/l	96
12) 1,1-Dichloroethene	2.319	96	867	0.539	ug/l #	70
27) cis-1,2-Dichloroethene	4.507	96	1398	0.648	ug/l	72
44) Trichloroethene	7.123	130	61461	31.038	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037252.D
Acq On : 25 Aug 2023 06:48
Operator : JC/MD
Sample : 04095-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-20230818

Quant Time: Aug 25 07:47:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

