

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037253.D
 Acq On : 25 Aug 2023 07:12
 Operator : JC/MD
 Sample : 04077-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20230817

Quant Time: Aug 25 07:47:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	157592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	264086	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99690	43.878	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.760%
35) Dibromofluoromethane	5.385	113	79772	45.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.220%
50) Toluene-d8	8.646	98	306889	48.373	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.740%
62) 4-Bromofluorobenzene	11.079	95	116683	46.801	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.600%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	35374	21.234	ug/l	98
12) 1,1-Dichloroethene	2.324	96	2357	1.453	ug/l #	63
27) cis-1,2-Dichloroethene	4.495	96	5144	2.366	ug/l	82
44) Trichloroethene	7.122	130	784736	388.338	ug/l	100
52) Toluene	8.720	92	1679	0.357	ug/l	95
64) Tetrachloroethene	9.274	164	8264	4.780	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037253.D
Acq On : 25 Aug 2023 07:12
Operator : JC/MD
Sample : 04077-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20230817

Quant Time: Aug 25 07:47:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

