

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037228.D
 Acq On : 24 Aug 2023 19:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 24 23:41:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	0.581	0.539	7.2	65	0.00
3 P	Chloromethane	0.508	0.639	-25.8#	82	0.00
4 C	Vinyl Chloride	0.553	0.589	-6.5#	74	0.00
5 T	Bromomethane	0.276	0.288	-4.3	75	0.00
6 T	Chloroethane	0.325	0.407	-25.2#	92	0.00
7 T	Trichlorofluoromethane	0.906	0.915	-1.0	73	0.00
8 T	Diethyl Ether	0.351	0.452	-28.8#	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.529	0.471	11.0	66	0.00
10 T	Methyl Iodide	0.642	0.645	-0.5	64	0.00
11 T	Tert butyl alcohol	0.169	0.222	-31.4#	87	0.00
12 CM	1,1-Dichloroethene	0.515	0.521	-1.2#	72	0.00
13 T	Acrolein	0.129	0.156	-20.9	81	0.00
14 T	Allyl chloride	0.797	0.930	-16.7	75	0.00
15 T	Acrylonitrile	0.302	0.406	-34.4#	87	0.00
16 T	Acetone	0.282	0.345	-22.3	83	0.00
17 T	Carbon Disulfide	1.309	1.126	14.0	62	0.00
18 T	Methyl Acetate	1.062	1.364	-28.4#	83	0.00
19 T	Methyl tert-butyl Ether	1.952	2.514	-28.8#	83	0.00
20 T	Methylene Chloride	0.631	0.763	-20.9	86	0.00
21 T	trans-1,2-Dichloroethene	0.585	0.622	-6.3	72	0.00
22 T	Diisopropyl ether	1.635	2.062	-26.1#	82	0.00
23 T	Vinyl Acetate	1.253	1.603	-27.9#	80	0.00
24 P	1,1-Dichloroethane	1.025	1.246	-21.6	81	0.00
25 T	2-Butanone	0.425	0.548	-28.9#	83	0.00
26 T	2,2-Dichloropropane	0.916	0.906	1.1	68	0.00
27 T	cis-1,2-Dichloroethene	0.690	0.833	-20.7	80	0.00
28 T	Bromochloromethane	0.458	0.591	-29.0#	83	0.00
29 T	Tetrahydrofuran	0.271	0.344	-26.9#	82	0.00
30 C	Chloroform	1.151	1.408	-22.3#	81	0.00
31 T	Cyclohexane	0.811	0.711	12.3	67	0.00
32 T	1,1,1-Trichloroethane	1.031	1.153	-11.8	77	0.00
33 S	1,2-Dichloroethane-d4	0.721	0.938	-30.1#	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.331	0.389	-17.5	87	0.00
36 T	1,1-Dichloropropene	0.470	0.411	12.6	70	0.00
37 T	Ethyl Acetate	0.481	0.532	-10.6	81	0.00
38 T	Carbon Tetrachloride	0.511	0.472	7.6	73	0.00
39 T	Methylcyclohexane	0.494	0.408	17.4	74	0.00
40 TM	Benzene	1.349	1.430	-6.0	79	0.00
41 T	Methacrylonitrile	0.250	0.285	-14.0	81	0.00
42 TM	1,2-Dichloroethane	0.529	0.594	-12.3	81	0.00
43 T	Isopropyl Acetate	0.763	0.865	-13.4	81	0.00
44 TM	Trichloroethene	0.383	0.391	-2.1	79	0.00
45 C	1,2-Dichloropropane	0.333	0.375	-12.6#	82	0.00
46 T	Dibromomethane	0.262	0.295	-12.6	81	0.00
47 T	Bromodichloromethane	0.502	0.576	-14.7	82	0.00
48 T	Methyl methacrylate	0.372	0.426	-14.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
 Data File : VX037228.D
 Acq On : 24 Aug 2023 19:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 24 23:41:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.013	-30.0#	93	0.00
50 S	Toluene-d8	1.201	1.334	-11.1	84	0.00
51 T	4-Methyl-2-Pentanone	0.463	0.558	-20.5	88	0.00
52 CM	Toluene	0.891	0.935	-4.9#	79	0.00
53 T	t-1,3-Dichloropropene	0.535	0.611	-14.2	79	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.643	-11.2	79	0.00
55 T	1,1,2-Trichloroethane	0.363	0.439	-20.9	88	0.00
56 T	Ethyl methacrylate	0.569	0.685	-20.4	86	0.00
57 T	1,3-Dichloropropane	0.612	0.715	-16.8	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.327	-19.8	87	0.00
59 T	2-Hexanone	0.353	0.440	-24.6	89	0.00
60 T	Dibromochloromethane	0.386	0.469	-21.5	84	0.00
61 T	1,2-Dibromoethane	0.397	0.465	-17.1	85	0.00
62 S	4-Bromofluorobenzene	0.472	0.552	-16.9	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.365	0.281	23.0	65	0.00
65 PM	Chlorobenzene	1.057	1.096	-3.7	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.437	-12.1	85	0.00
67 C	Ethyl Benzene	1.860	1.851	0.5#	81	0.00
68 T	m/p-Xylenes	0.718	0.718	0.0	81	0.00
69 T	o-Xylene	0.708	0.749	-5.8	84	0.00
70 T	Styrene	1.167	1.272	-9.0	84	0.00
71 P	Bromoform	0.292	0.360	-23.3	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.634	3.502	3.6	84	0.00
74 T	N-amyl acetate	1.412	1.606	-13.7	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.224	1.430	-16.8	95	0.00
76 T	1,2,3-Trichloropropane	1.065	1.197	-12.4	93	0.00
77 T	Bromobenzene	0.888	0.942	-6.1	87	0.00
78 T	n-propylbenzene	4.047	3.905	3.5	84	0.00
79 T	2-Chlorotoluene	2.586	2.643	-2.2	87	0.00
80 T	1,3,5-Trimethylbenzene	3.032	3.106	-2.4	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.381	0.416	-9.2	85	0.00
82 T	4-Chlorotoluene	2.985	3.034	-1.6	85	0.00
83 T	tert-Butylbenzene	2.891	3.049	-5.5	93	0.00
84 T	1,2,4-Trimethylbenzene	3.064	3.188	-4.0	88	0.00
85 T	sec-Butylbenzene	3.377	3.458	-2.4	92	0.00
86 T	p-Isopropyltoluene	2.907	3.022	-4.0	92	0.00
87 T	1,3-Dichlorobenzene	1.662	1.717	-3.3	87	0.00
88 T	1,4-Dichlorobenzene	1.717	1.731	-0.8	86	0.00
89 T	n-Butylbenzene	2.410	2.413	-0.1	90	0.00
90 T	Hexachloroethane	0.485	0.556	-14.6	96	0.00
91 T	1,2-Dichlorobenzene	1.627	1.808	-11.1	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.345	-21.5	91	0.00
93 T	1,2,4-Trichlorobenzene	0.952	1.003	-5.4	90	0.00
94 T	Hexachlorobutadiene	0.368	0.339	7.9	87	0.00
95 T	Naphthalene	4.171	4.025	3.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082423\
Data File : VX037228.D
Acq On : 24 Aug 2023 19:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 24 23:41:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.946	1.034	-9.3	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6