

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
 Data File : VX023932.D
 Acq On : 25 Aug 2021 12:46
 Operator : JC/MD
 Sample : M3433-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-EB-20210812

Quant Time: Aug 26 07:35:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	224388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100065	50.589	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.180%
35) Dibromofluoromethane	5.397	113	76916	47.874	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.740%
50) Toluene-d8	8.653	98	293560	48.007	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.020%
62) 4-Bromofluorobenzene	11.085	95	102478	45.524	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.040%
Target Compounds						
11) Tert butyl alcohol	2.971	59	21642	54.103	ug/l	98
16) Acetone	2.386	43	23787	26.896	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
Data File : VX023932.D
Acq On : 25 Aug 2021 12:46
Operator : JC/MD
Sample : M3433-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-EB-20210812

Quant Time: Aug 26 07:35:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

