

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
 Data File : VX023940.D
 Acq On : 25 Aug 2021 15:54
 Operator : JC/MD
 Sample : M3520-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-500-502

Quant Time: Aug 25 16:55:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	217076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95510	50.715	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.440%
35) Dibromofluoromethane	5.397	113	74721	48.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	8.653	98	278446	47.658	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.320%
62) 4-Bromofluorobenzene	11.085	95	96732	44.975	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%
Target Compounds						
						Qvalue
3) Chloromethane	1.294	50	609	0.390	ug/l	97
16) Acetone	2.392	43	4917	5.839	ug/l	98
17) Carbon Disulfide	2.514	76	2396	0.776	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
Data File : VX023940.D
Acq On : 25 Aug 2021 15:54
Operator : JC/MD
Sample : M3520-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-500-502

Quant Time: Aug 25 16:55:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

