

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
 Data File : VX023942.D
 Acq On : 25 Aug 2021 16:41
 Operator : JC/MD
 Sample : M3525-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-SW-17-GW-082021

Quant Time: Aug 26 07:37:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235282	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88294	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98351	51.250	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.500%	
35) Dibromofluoromethane	5.397	113	75823	48.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.380%	
50) Toluene-d8	8.653	98	283347	47.806	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.620%	
62) 4-Bromofluorobenzene	11.085	95	97330	44.608	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.220%	

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	764	0.527	ug/l	78
7) Trichlorofluoromethane	1.892	101	278815	113.675	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.337	101	1266	0.917	ug/l	98
12) 1,1-Dichloroethene	2.325	96	10526	7.977	ug/l	89
16) Acetone	2.392	43	3050	3.555	ug/l	89
27) cis-1,2-Dichloroethene	4.501	96	3408	2.032	ug/l	94
30) Chloroform	5.105	83	136007	47.652	ug/l	99
42) 1,2-Dichloroethane	6.098	62	2477	0.978	ug/l	90
44) Trichloroethene	7.135	130	178919	100.528	ug/l	98
64) Tetrachloroethene	9.281	164	6035	3.434	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
Data File : VX023942.D
Acq On : 25 Aug 2021 16:41
Operator : JC/MD
Sample : M3525-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-SW-17-GW-082021

Quant Time: Aug 26 07:37:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

