

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
 Data File : VX023945.D
 Acq On : 25 Aug 2021 17:52
 Operator : JC/MD
 Sample : M3525-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-SW8-GW-082121

Quant Time: Aug 26 07:37:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137741	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234853	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100476	52.265	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.540%
35) Dibromofluoromethane	5.397	113	77619	49.934	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.860%
50) Toluene-d8	8.653	98	288803	48.815	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.640%
62) 4-Bromofluorobenzene	11.085	95	101536	46.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.240%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.892	101	1194	0.486	ug/l	99
16) Acetone	2.386	43	2071	2.409	ug/l #	84
30) Chloroform	5.111	83	2673	0.935	ug/l	94
44) Trichloroethene	7.129	130	1639	0.923	ug/l	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
Data File : VX023945.D
Acq On : 25 Aug 2021 17:52
Operator : JC/MD
Sample : M3525-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-SW8-GW-082121

Quant Time: Aug 26 07:37:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

