

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
 Data File : VX023935.D
 Acq On : 25 Aug 2021 13:56
 Operator : JC/MD
 Sample : M3525-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-FB-082121

Quant Time: Aug 26 07:36:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.562	168	143700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247745	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98100	48.913	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.397	113	77402	47.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.400%
50) Toluene-d8	8.653	98	285331	45.719	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.440%#
62) 4-Bromofluorobenzene	11.085	95	96890	42.172	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.340%
Target Compounds						
16) Acetone	2.392	43	3294	3.673	ug/l	Qvalue 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082521\
Data File : VX023935.D
Acq On : 25 Aug 2021 13:56
Operator : JC/MD
Sample : M3525-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-FB-082121

Quant Time: Aug 26 07:36:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56




Instrument :
MSVOA_X
ClientSampleId :
WVLY-FB-082121


Tgt Ion:168 Resp: 143700
Ion Ratio Lower Upper
168 100
99 59.4 45.8 68.6



#16
Acetone
Concen: 3.673 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56

Tgt Ion: 43 Resp: 3294
Ion Ratio Lower Upper
43 100
58 30.0 27.8 41.8









#33

1,2-Dichloroethane-d4
Concen: 48.913 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56

Instrument :
MSVOA_X
ClientSampleId :
WVLY-FB-082121




Tgt Ion: 65 Resp: 98100
Ion Ratio Lower Upper
65 100
67 51.4 0.0 105.4




#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56

Tgt Ion: 114 Resp: 247745
Ion Ratio Lower Upper
114 100
63 21.4 0.0 40.8
88 16.1 0.0 32.8









#35
Dibromofluoromethane
Concen: 47.203 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56




Instrument :
MSVOA_X
ClientSampleId :
WVLY-FB-082121


Tgt Ion:113 Resp: 77402
Ion Ratio Lower Upper
113 100
111 102.8 81.9 122.9
192 19.7 16.7 25.1



#50
Toluene-d8
Concen: 45.719 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56

Tgt Ion: 98 Resp: 285331
Ion Ratio Lower Upper
98 100
100 63.9 51.9 77.9









#62
4-Bromofluorobenzene
Concen: 42.172 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56




Instrument :
MSVOA_X
ClientSampleId :
WVLY-FB-082121


Tgt Ion: 95 Resp: 96890
Ion Ratio Lower Upper
95 100
174 78.9 0.0 154.6
176 77.3 0.0 155.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX023935.D
Acq: 25 Aug 2021 13:56

Tgt Ion: 117 Resp: 228515
Ion Ratio Lower Upper
117 100
82 58.6 47.4 71.2
119 32.1 26.6 39.8





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.024 min Scan# 1794
 Delta R.T. 0.000 min
 Lab File: VX023935.D
 Acq: 25 Aug 2021 13:56

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-FB-082121

Tgt Ion:152 Resp: 90153
 Ion Ratio Lower Upper
 152 100
 115 60.0 41.1 123.3
 150 159.6 0.0 349.0

