

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030774.D
 Acq On : 25 Aug 2022 16:24
 Operator : JC/MD
 Sample : N4266-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Aug 26 02:40:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	235114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	391409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	348949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	137642	49.770	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.540%
35) Dibromofluoromethane	5.385	113	118275	49.168	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.340%
50) Toluene-d8	8.647	98	450452	53.737	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.480%
62) 4-Bromofluorobenzene	11.085	95	152685	48.925	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.840%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030774.D
Acq On : 25 Aug 2022 16:24
Operator : JC/MD
Sample : N4266-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Aug 26 02:40:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

