

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030777.D
 Acq On : 25 Aug 2022 17:35
 Operator : JC/MD
 Sample : N4222-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20220815

Quant Time: Aug 26 02:40:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	233236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	384671	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	343408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	138007	50.304	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.600%
35) Dibromofluoromethane	5.391	113	118169	49.984	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.960%
50) Toluene-d8	8.653	98	447346	54.302	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.600%
62) 4-Bromofluorobenzene	11.086	95	153092	49.915	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.820%
Target Compounds						
16) Acetone	2.386	43	2590	2.043	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030777.D
Acq On : 25 Aug 2022 17:35
Operator : JC/MD
Sample : N4222-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20220815

Quant Time: Aug 26 02:40:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

