

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030782.D
 Acq On : 25 Aug 2022 19:33
 Operator : JC/MD
 Sample : N4245-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D-20220816

Quant Time: Aug 26 02:41:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	224651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	384927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	333539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	134652	50.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.920%	
35) Dibromofluoromethane	5.391	113	117734	49.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.540%	
50) Toluene-d8	8.653	98	437159	53.030	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.060%	
62) 4-Bromofluorobenzene	11.085	95	148352	48.337	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.680%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	800	0.393	ug/l	86
19) Methyl tert-butyl Ether	3.117	73	6949	0.985	ug/l #	86
21) trans-1,2-Dichloroethene	3.093	96	741	0.339	ug/l #	78
27) cis-1,2-Dichloroethene	4.489	96	77199	28.566	ug/l	91
44) Trichloroethene	7.129	130	346711	124.515	ug/l	99
64) Tetrachloroethene	9.275	164	33987	12.936	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030782.D
Acq On : 25 Aug 2022 19:33
Operator : JC/MD
Sample : N4245-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D-20220816

Quant Time: Aug 26 02:41:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

