

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030794.D
 Acq On : 26 Aug 2022 01:03
 Operator : JC/MD
 Sample : N4245-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D1-20220816

Quant Time: Aug 26 02:44:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	228434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	381436	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	325489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138485	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	136147	50.669	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.340%
35) Dibromofluoromethane	5.397	113	120016	51.196	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.400%
50) Toluene-d8	8.653	98	432834	52.986	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.980%
62) 4-Bromofluorobenzene	11.085	95	143330	47.128	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.260%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2483	2.000	ug/l	# 87
29) Tetrahydrofuran	5.013	42	16674	12.773	ug/l	99
44) Trichloroethene	7.135	130	796	0.288	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030794.D
Acq On : 26 Aug 2022 01:03
Operator : JC/MD
Sample : N4245-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D1-20220816

Quant Time: Aug 26 02:44:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

