

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030796.D
 Acq On : 26 Aug 2022 01:50
 Operator : JC/MD
 Sample : N4245-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D1-20220816

Quant Time: Aug 26 02:45:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	233247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	395521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	339229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139692	50.915	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.840%	
35) Dibromofluoromethane	5.397	113	121793	50.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.200%	
50) Toluene-d8	8.653	98	444575	52.485	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.960%	
62) 4-Bromofluorobenzene	11.085	95	144235	45.737	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.480%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	2559	1.127	ug/l	87
12) 1,1-Dichloroethene	2.319	96	3034	1.435	ug/l #	85
16) Acetone	2.386	43	2976	2.347	ug/l	94
27) cis-1,2-Dichloroethene	4.501	96	3128	1.115	ug/l	91
30) Chloroform	5.117	83	1626	0.347	ug/l	99
38) Carbon Tetrachloride	5.684	117	4805	1.376	ug/l	87
44) Trichloroethene	7.129	130	284555	99.456	ug/l	99
55) 1,1,2-Trichloroethane	9.159	97	1624	0.591	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030796.D
Acq On : 26 Aug 2022 01:50
Operator : JC/MD
Sample : N4245-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D1-20220816

Quant Time: Aug 26 02:45:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

