

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030801.D
 Acq On : 26 Aug 2022 03:47
 Operator : JC/MD
 Sample : N4262-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102I

Quant Time: Aug 26 04:57:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	234966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	393389	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	351708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	208663	75.498	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 151.000%#	
35) Dibromofluoromethane	5.391	113	181299	74.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 149.980%#	
50) Toluene-d8	8.653	98	666013	79.053	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 158.100%#	
62) 4-Bromofluorobenzene	11.085	95	226370	72.171	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 144.340%#	
Target Compounds						
					Qvalue	
21) trans-1,2-Dichloroethene	3.099	96	1426	0.623	ug/l	86
27) cis-1,2-Dichloroethene	4.495	96	7545	2.669	ug/l	92
30) Chloroform	5.099	83	3019	0.639	ug/l	86
44) Trichloroethene	7.129	130	50487	17.742	ug/l	97
64) Tetrachloroethene	9.275	164	89948	32.466	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030801.D
Acq On : 26 Aug 2022 03:47
Operator : JC/MD
Sample : N4262-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102I

Quant Time: Aug 26 04:57:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

