

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030803.D
 Acq On : 26 Aug 2022 04:34
 Operator : JC/MD
 Sample : N4262-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-101I

Quant Time: Aug 26 05:38:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	235450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	391722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	352751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	212525	76.737	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 153.480%#	
35) Dibromofluoromethane	5.391	113	182602	75.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 151.700%#	
50) Toluene-d8	8.652	98	668999	79.745	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 159.500%#	
62) 4-Bromofluorobenzene	11.085	95	226382	72.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 144.960%#	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	39609	18.558	ug/l	95
24) 1,1-Dichloroethane	3.617	63	78534	17.467	ug/l	100
27) cis-1,2-Dichloroethene	4.495	96	57284	20.225	ug/l	94
30) Chloroform	5.086	83	2335	0.493	ug/l	89
32) 1,1,1-Trichloroethane	5.385	97	6178	1.615	ug/l #	49
44) Trichloroethene	7.128	130	53500	18.880	ug/l	95
49) 1,4-Dioxane	7.665	88	4575	63.565	ug/l	96
64) Tetrachloroethene	9.274	164	18861	6.788	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030803.D
Acq On : 26 Aug 2022 04:34
Operator : JC/MD
Sample : N4262-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-101I

Quant Time: Aug 26 05:38:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

