

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030805.D
 Acq On : 26 Aug 2022 05:21
 Operator : JC/MD
 Sample : N4262-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-107(42.5FT)

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/26/2022
 Supervised By :Mahesh Dadoda 08/26/2022

Quant Time: Aug 26 05:57:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	234512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	389582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	350109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	206374	74.814	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 149.620%#			
35) Dibromofluoromethane	5.391	113	177352	74.072	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 148.140%#			
50) Toluene-d8	8.653	98	648053	77.673	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 155.340%#			
62) 4-Bromofluorobenzene	11.085	95	216920	69.834	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 139.660%#			
Target Compounds						Qvalue
21) trans-1,2-Dichloroethene	3.099	96	1192	0.522	ug/l #	77
27) cis-1,2-Dichloroethene	4.489	96	7012	2.486	ug/l	97
30) Chloroform	5.111	83	3046m	0.646	ug/l	
44) Trichloroethene	7.135	130	45875	16.278	ug/l	97
64) Tetrachloroethene	9.275	164	80421	29.160	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030805.D
Acq On : 26 Aug 2022 05:21
Operator : JC/MD
Sample : N4262-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-107(42.5FT)

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/26/2022
Supervised By :Mahesh Dadoda 08/26/2022

Quant Time: Aug 26 05:57:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

