

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030807.D
 Acq On : 26 Aug 2022 06:08
 Operator : JC/MD
 Sample : N4278-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20220818

Quant Time: Aug 26 07:49:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	236643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	398993	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	356372	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	216006	77.601	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 155.200%#	
35) Dibromofluoromethane	5.391	113	187385	76.417	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 152.840%#	
50) Toluene-d8	8.653	98	693296	81.136	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 162.280%#	
62) 4-Bromofluorobenzene	11.079	95	230338	72.404	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 144.800%#	
Target Compounds						
16) Acetone	2.386	43	2626	2.042	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030807.D
Acq On : 26 Aug 2022 06:08
Operator : JC/MD
Sample : N4278-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20220818

Quant Time: Aug 26 07:49:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

