

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030784.D
 Acq On : 25 Aug 2022 20:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 02:42:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.463	0.448	3.2	93	0.00
3 P	Chloromethane	0.432	0.380	12.0	94	0.00
4 C	Vinyl Chloride	0.508	0.452	11.0#	93	0.00
5 T	Bromomethane	0.233	0.199	14.6	90	0.00
6 T	Chloroethane	0.313	0.280	10.5	98	0.00
7 T	Trichlorofluoromethane	0.765	0.697	8.9	93	0.00
8 T	Diethyl Ether	0.290	0.283	2.4	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.487	0.447	8.2	92	0.00
10 T	Methyl Iodide	0.337	0.340	-0.9	82	0.00
11 T	Tert butyl alcohol	0.128	0.147	-14.8	121	-0.02
12 CM	1,1-Dichloroethene	0.453	0.418	7.7#	95	0.00
13 T	Acrolein	0.097	0.103	-6.2	88	0.00
14 T	Allyl chloride	0.839	0.777	7.4	93	0.00
15 T	Acrylonitrile	0.313	0.330	-5.4	102	0.00
16 T	Acetone	0.272	0.275	-1.1	100	0.00
17 T	Carbon Disulfide	1.052	0.868	17.5	92	0.00
18 T	Methyl Acetate	0.835	0.849	-1.7	100	0.00
19 T	Methyl tert-butyl Ether	1.569	1.565	0.3	96	0.00
20 T	Methylene Chloride	0.559	0.564	-0.9	106	0.00
21 T	trans-1,2-Dichloroethene	0.487	0.440	9.7	93	0.00
22 T	Diisopropyl ether	1.685	1.661	1.4	96	0.00
23 T	Vinyl Acetate	1.351	1.360	-0.7	96	0.00
24 P	1,1-Dichloroethane	0.955	0.917	4.0	96	0.00
25 T	2-Butanone	0.438	0.455	-3.9	101	0.00
26 T	2,2-Dichloropropane	0.692	0.504	27.2#	69	0.00
27 T	cis-1,2-Dichloroethene	0.601	0.572	4.8	96	0.00
28 T	Bromochloromethane	0.411	0.391	4.9	94	0.00
29 T	Tetrahydrofuran	0.286	0.297	-3.8	102	0.00
30 C	Chloroform	1.005	0.958	4.7#	97	0.00
31 T	Cyclohexane	0.752	0.704	6.4	91	0.00
32 T	1,1,1-Trichloroethane	0.813	0.794	2.3	95	0.00
33 S	1,2-Dichloroethane-d4	0.588	0.561	4.6	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.307	0.294	4.2	97	0.00
36 T	1,1-Dichloropropene	0.418	0.370	11.5	92	0.00
37 T	Ethyl Acetate	0.496	0.490	1.2	101	0.00
38 T	Carbon Tetrachloride	0.441	0.408	7.5	94	0.00
39 T	Methylcyclohexane	0.501	0.441	12.0	90	0.00
40 TM	Benzene	1.252	1.199	4.2	98	0.00
41 T	Methacrylonitrile	0.277	0.275	0.7	98	0.00
42 TM	1,2-Dichloroethane	0.444	0.430	3.2	99	0.00
43 T	Isopropyl Acetate	0.747	0.760	-1.7	100	0.00
44 TM	Trichloroethene	0.362	0.348	3.9	98	0.00
45 C	1,2-Dichloropropane	0.346	0.340	1.7#	98	0.00
46 T	Dibromomethane	0.246	0.232	5.7	96	0.00
47 T	Bromodichloromethane	0.463	0.462	0.2	97	0.00
48 T	Methyl methacrylate	0.372	0.387	-4.0	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
 Data File : VX030784.D
 Acq On : 25 Aug 2022 20:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 02:42:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.011	-22.2	117	-0.01
50 S	Toluene-d8	1.071	1.015	5.2	96	0.00
51 T	4-Methyl-2-Pentanone	0.504	0.519	-3.0	101	0.00
52 CM	Toluene	0.781	0.748	4.2#	95	0.00
53 T	t-1,3-Dichloropropene	0.448	0.433	3.3	91	0.00
54 T	cis-1,3-Dichloropropene	0.504	0.485	3.8	92	0.00
55 T	1,1,2-Trichloroethane	0.347	0.341	1.7	96	0.00
56 T	Ethyl methacrylate	0.498	0.526	-5.6	98	0.00
57 T	1,3-Dichloropropane	0.560	0.541	3.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.274	-3.0	95	0.00
59 T	2-Hexanone	0.378	0.407	-7.7	103	0.00
60 T	Dibromochloromethane	0.362	0.351	3.0	93	0.00
61 T	1,2-Dibromoethane	0.352	0.346	1.7	96	0.00
62 S	4-Bromofluorobenzene	0.399	0.398	0.3	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.394	0.363	7.9	94	0.00
65 PM	Chlorobenzene	0.967	0.899	7.0	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.350	3.6	95	0.00
67 C	Ethyl Benzene	1.671	1.596	4.5#	96	0.00
68 T	m/p-Xylenes	0.640	0.622	2.8	97	0.00
69 T	o-Xylene	0.634	0.623	1.7	97	0.00
70 T	Styrene	1.051	1.061	-1.0	96	0.00
71 P	Bromoform	0.300	0.290	3.3	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.351	3.193	4.7	95	0.00
74 T	N-amyl acetate	1.305	1.363	-4.4	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.146	1.117	2.5	100	0.00
76 T	1,2,3-Trichloropropane	1.095	1.043	4.7	86	0.00
77 T	Bromobenzene	0.826	0.777	5.9	97	0.00
78 T	n-propylbenzene	3.854	3.706	3.8	95	0.00
79 T	2-Chlorotoluene	2.361	2.243	5.0	98	0.00
80 T	1,3,5-Trimethylbenzene	2.804	2.692	4.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.325	4.4	91	0.00
82 T	4-Chlorotoluene	2.646	2.536	4.2	95	0.00
83 T	tert-Butylbenzene	2.699	2.588	4.1	92	0.00
84 T	1,2,4-Trimethylbenzene	2.716	2.701	0.6	95	0.00
85 T	sec-Butylbenzene	3.517	3.483	1.0	94	0.00
86 T	p-Isopropyltoluene	2.892	2.846	1.6	94	0.00
87 T	1,3-Dichlorobenzene	1.557	1.491	4.2	90	0.00
88 T	1,4-Dichlorobenzene	1.598	1.511	5.4	92	0.00
89 T	n-Butylbenzene	2.678	2.516	6.0	85	0.00
90 T	Hexachloroethane	0.561	0.502	10.5	86	0.00
91 T	1,2-Dichlorobenzene	1.639	1.548	5.6	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.276	6.4	91	0.00
93 T	1,2,4-Trichlorobenzene	1.025	0.950	7.3	85	0.00
94 T	Hexachlorobutadiene	0.427	0.390	8.7	86	0.00
95 T	Naphthalene	3.569	3.566	0.1	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082522\
Data File : VX030784.D
Acq On : 25 Aug 2022 20:20
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 26 02:42:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.084	0.998	7.9	83	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6