

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082523\
 Data File : VX037267.D
 Acq On : 25 Aug 2023 13:32
 Operator : JC/MD
 Sample : 04025-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP06-20230814

Quant Time: Aug 26 02:20:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240586	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91599	43.545	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.080%
35) Dibromofluoromethane	5.385	113	73229	45.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.647	98	280779	48.580	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	102209	45.000	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.000%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1542	1.000	ug/l	95
12) 1,1-Dichloroethene	2.319	96	386	0.257	ug/l	85
27) cis-1,2-Dichloroethene	4.495	96	1931	0.959	ug/l	84
44) Trichloroethene	7.129	130	349534	189.868	ug/l	94
64) Tetrachloroethene	9.275	164	2096	1.341	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082523\
Data File : VX037267.D
Acq On : 25 Aug 2023 13:32
Operator : JC/MD
Sample : 04025-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP06-20230814

Quant Time: Aug 26 02:20:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

