

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082523\
 Data File : VX037270.D
 Acq On : 25 Aug 2023 14:44
 Operator : JC/MD
 Sample : 04095-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20230818

Quant Time: Aug 26 02:21:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247903	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91912	42.668	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.340%
35) Dibromofluoromethane	5.385	113	75489	45.979	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.960%
50) Toluene-d8	8.647	98	289381	48.591	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	106153	45.357	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.720%
Target Compounds						
30) Chloroform	5.093	83	4024	1.170	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082523\
Data File : VX037270.D
Acq On : 25 Aug 2023 14:44
Operator : JC/MD
Sample : 04095-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20230818

Quant Time: Aug 26 02:21:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

