

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082523\
 Data File : VX037279.D
 Acq On : 25 Aug 2023 18:22
 Operator : JC/MD
 Sample : 04095-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT17411-20230818

Quant Time: Aug 26 02:24:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	85858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84556	43.216	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.440%
35) Dibromofluoromethane	5.391	113	68662	46.642	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	8.647	98	260728	48.827	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.660%
62) 4-Bromofluorobenzene	11.079	95	92477	44.068	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.140%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1275	0.913	ug/l	86
24) 1,1-Dichloroethane	3.611	63	8364	3.006	ug/l	97
44) Trichloroethene	7.129	130	3465	2.037	ug/l	96
64) Tetrachloroethene	9.269	164	506	0.350	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082523\
Data File : VX037279.D
Acq On : 25 Aug 2023 18:22
Operator : JC/MD
Sample : 04095-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT17411-20230818

Quant Time: Aug 26 02:24:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

