

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082523\  
 Data File : VX037262.D  
 Acq On : 25 Aug 2023 11:31  
 Operator : JC/MD  
 Sample : VX0825WBSD01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0825WBSD01

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
 Supervised By :Semsettin Yesilyurt 08/28/2023

Quant Time: Aug 26 02:17:55 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081723W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 18 09:30:38 2023  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.549          | 168  | 137154              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 233448              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 218186              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 113465              | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.952          | 65   | 102149              | 51.660  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 103.320% |         |        |          |
| 35) Dibromofluoromethane     | 5.385          | 113  | 83533               | 54.029  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 108.060% |         |        |          |
| 50) Toluene-d8               | 8.646          | 98   | 295162              | 52.631  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 105.260% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 118938              | 53.966  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 107.940% |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 31470               | 19.753  | ug/l   | 98       |
| 3) Chloromethane             | 1.300          | 50   | 29217               | 20.962  | ug/l   | 97       |
| 4) Vinyl Chloride            | 1.373          | 62   | 29462               | 19.419  | ug/l   | 98       |
| 5) Bromomethane              | 1.599          | 94   | 21082               | 27.896  | ug/l   | 97       |
| 6) Chloroethane              | 1.684          | 64   | 20457               | 22.914  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.885          | 101  | 50463               | 20.305  | ug/l   | 99       |
| 8) Diethyl Ether             | 2.129          | 74   | 20369               | 21.184  | ug/l   | 75       |
| 9) 1,1,2-Trichlorotrifluo... | 2.324          | 101  | 29292               | 20.204  | ug/l   | 97       |
| 10) Methyl Iodide            | 2.452          | 142  | 33212               | 18.848  | ug/l   | 95       |
| 11) Tert butyl alcohol       | 2.958          | 59   | 45128               | 97.293  | ug/l # | 92       |
| 12) 1,1-Dichloroethene       | 2.318          | 96   | 26386               | 18.689  | ug/l   | 86       |
| 13) Acrolein                 | 2.233          | 56   | 35633               | 100.984 | ug/l   | 99       |
| 14) Allyl chloride           | 2.660          | 41   | 42583               | 19.486  | ug/l # | 91       |
| 15) Acrylonitrile            | 3.062          | 53   | 89865               | 108.548 | ug/l   | 98       |
| 16) Acetone                  | 2.379          | 43   | 82462               | 106.602 | ug/l # | 89       |
| 17) Carbon Disulfide         | 2.513          | 76   | 55559               | 15.471  | ug/l   | 98       |
| 18) Methyl Acetate           | 2.702          | 43   | 59773               | 20.523  | ug/l # | 87       |
| 19) Methyl tert-butyl Ether  | 3.111          | 73   | 110686              | 20.671  | ug/l   | 96       |
| 20) Methylene Chloride       | 2.788          | 84   | 35014               | 20.241  | ug/l   | 86       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 29756               | 18.540  | ug/l   | 86       |
| 22) Diisopropyl ether        | 3.757          | 45   | 93422               | 20.833  | ug/l # | 96       |
| 23) Vinyl Acetate            | 3.720          | 43   | 351580              | 102.310 | ug/l # | 91       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 56649               | 20.145  | ug/l   | 96       |
| 25) 2-Butanone               | 4.556          | 43   | 121656              | 104.450 | ug/l # | 86       |
| 26) 2,2-Dichloropropane      | 4.476          | 77   | 48182               | 19.185  | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 37278               | 19.702  | ug/l   | 91       |
| 28) Bromochloromethane       | 4.897          | 49   | 34363               | 27.354  | ug/l # | 78       |
| 29) Tetrahydrofuran          | 5.007          | 42   | 74975               | 100.938 | ug/l # | 83       |
| 30) Chloroform               | 5.092          | 83   | 63631               | 20.157  | ug/l   | 100      |
| 31) Cyclohexane              | 5.470          | 56   | 42385               | 19.064  | ug/l # | 84       |
| 32) 1,1,1-Trichloroethane    | 5.379          | 97   | 54600               | 19.301  | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 5.690          | 75   | 41657               | 18.965  | ug/l   | 97       |
| 37) Ethyl Acetate            | 4.714          | 43   | 45450               | 20.256  | ug/l # | 93       |
| 38) Carbon Tetrachloride     | 5.677          | 117  | 47097               | 19.738  | ug/l   | 95       |
| 39) Methylcyclohexane        | 7.378          | 83   | 51220               | 22.194  | ug/l   | 92       |
| 40) Benzene                  | 6.037          | 78   | 128070              | 20.330  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082523\  
 Data File : VX037262.D  
 Acq On : 25 Aug 2023 11:31  
 Operator : JC/MD  
 Sample : VX0825WBSD01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0825WBSD01

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
 Supervised By :Semsettin Yesilyurt 08/28/2023

Quant Time: Aug 26 02:17:55 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081723W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 18 09:30:38 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.915  | 41   | 24768    | 21.234  | ug/l # | 88       |
| 42) 1,2-Dichloroethane        | 6.086  | 62   | 51470    | 20.834  | ug/l   | 95       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 72987    | 20.489  | ug/l # | 91       |
| 44) Trichloroethene           | 7.122  | 130  | 34505    | 19.316  | ug/l   | 97       |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 33317    | 21.449  | ug/l   | 97       |
| 46) Dibromomethane            | 7.580  | 93   | 25881    | 21.150  | ug/l   | 94       |
| 47) Bromodichloromethane      | 7.823  | 83   | 48480    | 20.684  | ug/l   | 95       |
| 48) Methyl methacrylate       | 7.689  | 41   | 35844    | 20.633  | ug/l # | 86       |
| 49) 1,4-Dioxane               | 7.659  | 88   | 19927    | 425.907 | ug/l # | 86       |
| 51) 4-Methyl-2-Pentanone      | 8.573  | 43   | 234713   | 108.507 | ug/l   | 90       |
| 52) Toluene                   | 8.720  | 92   | 84733    | 20.357  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.976  | 75   | 50095    | 20.073  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 53839    | 19.964  | ug/l # | 88       |
| 55) 1,1,2-Trichloroethane     | 9.152  | 97   | 37473    | 22.109  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 56247    | 21.187  | ug/l   | 87       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 62153    | 21.753  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.238  | 63   | 140763   | 110.487 | ug/l   | 92       |
| 59) 2-Hexanone                | 9.427  | 43   | 182782   | 110.822 | ug/l   | 89       |
| 60) Dibromochloromethane      | 9.518  | 129  | 38038    | 21.079  | ug/l   | 98       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 39921    | 21.513  | ug/l   | 100      |
| 64) Tetrachloroethene         | 9.274  | 164  | 27685    | 17.393  | ug/l   | 99       |
| 65) Chlorobenzene             | 10.079 | 112  | 95172    | 20.638  | ug/l   | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 10.164 | 131  | 36017    | 21.162  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 164210   | 20.237  | ug/l   | 98       |
| 68) m/p-Xylenes               | 10.299 | 106  | 128203   | 40.913  | ug/l   | 100      |
| 69) o-Xylene                  | 10.640 | 106  | 64211    | 20.769  | ug/l   | 99       |
| 70) Styrene                   | 10.658 | 104  | 108210   | 21.248  | ug/l   | 97       |
| 71) Bromoform                 | 10.799 | 173  | 26191    | 20.573  | ug/l # | 100      |
| 73) Isopropylbenzene          | 10.963 | 105  | 169670   | 20.573  | ug/l   | 99       |
| 74) N-amyl acetate            | 10.841 | 43   | 64374    | 20.088  | ug/l # | 90       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 60861    | 21.913  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.237 | 75   | 49281m   | 20.394  | ug/l   |          |
| 77) Bromobenzene              | 11.195 | 156  | 40448    | 20.082  | ug/l   | 93       |
| 78) n-propylbenzene           | 11.305 | 91   | 192791   | 20.993  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.365 | 91   | 120584   | 20.547  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 147168   | 21.392  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 15592    | 18.022  | ug/l   | 93       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 139258   | 20.557  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.713 | 119  | 146528   | 22.333  | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 147317   | 21.188  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 178787   | 23.329  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.006 | 119  | 151075   | 22.900  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 78785    | 20.892  | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 78975    | 20.271  | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 125884   | 23.015  | ug/l   | 99       |
| 90) Hexachloroethane          | 12.536 | 117  | 24780    | 22.504  | ug/l   | 93       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 79183    | 21.444  | ug/l   | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.938 | 75   | 14059    | 21.785  | ug/l   | 89       |
| 93) 1,2,4-Trichlorobenzene    | 13.585 | 180  | 48373    | 22.388  | ug/l   | 97       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 20506    | 24.567  | ug/l   | 98       |
| 95) Naphthalene               | 13.774 | 128  | 172408   | 18.216  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 48089    | 22.402  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX082523\  
Data File : VX037262.D  
Acq On : 25 Aug 2023 11:31  
Operator : JC/MD  
Sample : VX0825WBSD01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023

Quant Time: Aug 26 02:17:55 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081723W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 18 09:30:38 2023  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.549 min Scan# 71  
Delta R.T. -0.001 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion:168 Resp: 137154  
Ion Ratio Lower Upper  
168 100  
99 63.9 56.6 85.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#2  
Dichlorodifluoromethane  
Concen: 19.753 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion: 85 Resp: 31470  
Ion Ratio Lower Upper  
85 100  
87 32.0 16.7 50.1





#3  
Chloromethane  
Concen: 20.962 ug/l  
RT: 1.300 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 50 Resp: 2921  
Ion Ratio Lower Upper  
50 100  
52 33.7 25.5 38.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#4  
Vinyl Chloride  
Concen: 19.419 ug/l  
RT: 1.373 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 62 Resp: 29462  
Ion Ratio Lower Upper  
62 100  
64 32.6 25.4 38.0





#5  
Bromomethane  
Concen: 27.896 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 94 Resp: 2108  
Ion Ratio Lower Upper  
94 100  
96 96.3 74.6 111.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#6  
Chloroethane  
Concen: 22.914 ug/l  
RT: 1.684 min Scan# 98  
Delta R.T. 0.012 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31



Tgt Ion: 64 Resp: 20457  
Ion Ratio Lower Upper  
64 100  
66 30.8 25.1 37.7









#7  
Trichlorofluoromethane  
Concen: 20.305 ug/l  
RT: 1.885 min Scan# 11  
Delta R.T. 0.005 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01

Tgt Ion:101 Resp: 5046  
Ion Ratio Lower Upper  
101 100  
103 65.3 51.4 77.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#8  
Diethyl Ether  
Concen: 21.184 ug/l  
RT: 2.129 min Scan# 171  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Tgt Ion: 74 Resp: 20369  
Ion Ratio Lower Upper  
74 100  
45 80.9 53.3 159.8









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.204 ug/l  
RT: 2.324 min Scan# 203  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01

Tgt Ion:101 Resp: 2929  
Ion Ratio Lower Upper  
101 100  
85 43.9 37.0 55.6  
151 75.3 58.6 87.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#10  
Methyl Iodide  
Concen: 18.848 ug/l  
RT: 2.452 min Scan# 224  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion:142 Resp: 33212  
Ion Ratio Lower Upper  
142 100  
127 47.7 41.6 62.4  
141 13.9 11.6 17.4





#11  
Tert butyl alcohol  
Concen: 97.293 ug/l  
RT: 2.958 min Scan# 30  
Delta R.T. -0.019 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 59 Resp: 45128  
Ion Ratio Lower Upper  
59 100  
57 10.2 6.0 9.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#12  
1,1-Dichloroethene  
Concen: 18.689 ug/l  
RT: 2.318 min Scan# 202  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 96 Resp: 26386  
Ion Ratio Lower Upper  
96 100  
61 157.6 144.6 216.8  
98 67.5 50.4 75.6





#13  
Acrolein  
Concen: 100.984 ug/l  
RT: 2.233 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 56 Resp: 3563  
Ion Ratio Lower Upper  
56 100  
55 70.9 57.1 85.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#14  
Allyl chloride  
Concen: 19.486 ug/l  
RT: 2.660 min Scan# 258  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 41 Resp: 42583  
Ion Ratio Lower Upper  
41 100  
39 65.8 55.8 83.6  
76 41.3 24.3 36.5





#15  
Acrylonitrile  
Concen: 108.548 ug/l  
RT: 3.062 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion: 53 Resp: 8986  
Ion Ratio Lower Upper  
53 100  
52 80.5 66.0 99.0  
51 35.7 29.8 44.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#16  
Acetone  
Concen: 106.602 ug/l  
RT: 2.379 min Scan# 212  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Tgt Ion: 43 Resp: 82462  
Ion Ratio Lower Upper  
43 100  
58 33.1 21.9 32.9#









#17  
Carbon Disulfide  
Concen: 15.471 ug/l  
RT: 2.513 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01

Tgt Ion: 76 Resp: 5555  
Ion Ratio Lower Upper  
76 100  
78 9.4 7.0 10.6




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#18  
Methyl Acetate  
Concen: 20.523 ug/l  
RT: 2.702 min Scan# 265  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion: 43 Resp: 59773  
Ion Ratio Lower Upper  
43 100  
74 28.4 17.8 26.8#





#19  
Methyl tert-butyl Ether  
Concen: 20.671 ug/l  
RT: 3.111 min Scan# 311  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion: 73 Resp: 110680  
Ion Ratio Lower Upper  
73 100  
57 19.7 17.3 25.9




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#20  
Methylene Chloride  
Concen: 20.241 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion: 84 Resp: 35014  
Ion Ratio Lower Upper  
84 100  
49 106.9 103.6 155.4  
51 34.5 33.8 50.6  
86 65.5 49.4 74.0





#21  
trans-1,2-Dichloroethene  
Concen: 18.540 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 96 Resp: 29750  
Ion Ratio Lower Upper  
96 100  
61 125.7 119.2 178.8  
98 64.3 48.7 73.1

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#22  
Diisopropyl ether  
Concen: 20.833 ug/l  
RT: 3.757 min Scan# 438  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 45 Resp: 93422  
Ion Ratio Lower Upper  
45 100  
43 86.3 69.7 104.5  
87 33.5 20.8 31.2#  
59 12.8 9.0 13.4





#23  
 Vinyl Acetate  
 Concen: 102.310 ug/l  
 RT: 3.720 min Scan# 41  
 Delta R.T. -0.000 min  
 Lab File: VX037262.D  
 Acq: 25 Aug 2023 11:31




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0825WBSD01



Tgt Ion: 43 Resp: 351580  
 Ion Ratio Lower Upper  
 43 100  
 86 13.0 7.8 11.8




Manual Integrations  
 APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
 Supervised By :Semsettin Yesilyurt 08/28/2023



#24  
 1,1-Dichloroethane  
 Concen: 20.145 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. 0.006 min  
 Lab File: VX037262.D  
 Acq: 25 Aug 2023 11:31

Tgt Ion: 63 Resp: 56649  
 Ion Ratio Lower Upper  
 63 100  
 98 8.2 3.3 9.8  
 100 5.2 2.1 6.5





#25

2-Butanone

Concen: 104.450 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01

Tgt Ion: 43 Resp: 121650

Ion Ratio Lower Upper

43 100





72 31.2 19.4 29.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023

Supervised By :Semsettin Yesilyurt 08/28/2023



#26

2,2-Dichloropropane

Concen: 19.185 ug/l

RT: 4.476 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX037262.D




Acq: 25 Aug 2023 11:31

Tgt Ion: 77 Resp: 48182

Ion Ratio Lower Upper

77 100

97 22.5 10.8 32.3





#27  
cis-1,2-Dichloroethene  
Concen: 19.702 ug/l  
RT: 4.489 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :





VX0825WBSD01

Tgt Ion: 96 Resp: 37278  
Ion Ratio Lower Upper  
96 100  
61 139.3 0.0 312.2  
98 65.2 0.0 129.0

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#28  
Bromochloromethane  
Concen: 27.354 ug/l  
RT: 4.897 min Scan# 625  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion: 49 Resp: 34363  
Ion Ratio Lower Upper  
49 100  
129 2.3 0.0 3.8  
130 90.6 57.5 86.3#





#29  
Tetrahydrofuran  
Concen: 100.938 ug/l  
RT: 5.007 min Scan# 643  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 42 Resp: 74975  
Ion Ratio Lower Upper  
42 100  
72 54.8 34.2 51.4  
71 49.7 32.2 48.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#30  
Chloroform  
Concen: 20.157 ug/l  
RT: 5.092 min Scan# 657  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 83 Resp: 63631  
Ion Ratio Lower Upper  
83 100  
85 63.0 50.3 75.5





#31  
Cyclohexane  
Concen: 19.064 ug/l  
RT: 5.470 min Scan# 719  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 56 Resp: 42385  
Ion Ratio Lower Upper  
56 100  
69 38.8 23.4 35.2  
84 95.9 65.9 98.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#32  
1,1,1-Trichloroethane  
Concen: 19.301 ug/l  
RT: 5.379 min Scan# 704  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 97 Resp: 54600  
Ion Ratio Lower Upper  
97 100  
99 65.3 51.1 76.7  
61 42.1 36.4 54.6





#33  
1,2-Dichloroethane-d4  
Concen: 51.660 ug/l  
RT: 5.952 min Scan# 799  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 65 Resp: 102149  
Ion Ratio Lower Upper  
65 100  
67 50.8 0.0 100.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.763 min Scan# 931  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion:114 Resp: 233448  
Ion Ratio Lower Upper  
114 100  
63 19.5 0.0 45.0  
88 15.9 0.0 36.6





#35  
Dibromofluoromethane  
Concen: 54.029 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 113 Resp: 8353  
Ion Ratio Lower Upper  
113 100  
111 101.9 82.9 124.3  
192 18.3 13.3 19.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#36  
1,1-Dichloropropene  
Concen: 18.965 ug/l  
RT: 5.690 min Scan# 755  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 75 Resp: 41657  
Ion Ratio Lower Upper  
75 100  
110 36.4 17.0 50.9  
77 31.2 24.7 37.1





#37  
Ethyl Acetate  
Concen: 20.256 ug/l  
RT: 4.714 min Scan# 595  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion: 43 Resp: 45450  
Ion Ratio Lower Upper  
43 100  
61 15.2 10.6 16.0  
70 14.0 8.3 12.5




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#38  
Carbon Tetrachloride  
Concen: 19.738 ug/l  
RT: 5.677 min Scan# 753  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Tgt Ion:117 Resp: 47097  
Ion Ratio Lower Upper  
117 100  
119 92.1 78.2 117.2  
121 29.7 24.4 36.6





#39

Methylcyclohexane

Concen: 22.194 ug/l

RT: 7.378 min Scan# 1032

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01





Tgt Ion: 83 Resp: 51220

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 55  | 71.7  | 64.6  | 97.0  |
| 98  | 48.2  | 36.7  | 55.1  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#40

Benzene

Concen: 20.330 ug/l

RT: 6.037 min Scan# 812




Delta R.T. -0.000 min


Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Tgt Ion: 78 Resp: 128070


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 78  | 100   |       |       |
| 77  | 23.6  | 19.3  | 28.9  |





#41  
Methacrylonitrile  
Concen: 21.234 ug/l  
RT: 4.915 min Scan# 61  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion: 41 Resp: 2476  
Ion Ratio Lower Upper  
41 100  
39 56.1 45.2 67.8  
67 83.7 52.6 78.8  
52 33.6 22.7 34.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#42  
1,2-Dichloroethane  
Concen: 20.834 ug/l  
RT: 6.086 min Scan# 820  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion: 62 Resp: 51470  
Ion Ratio Lower Upper  
62 100  
98 9.7 0.0 16.0





#43  
Isopropyl Acetate  
Concen: 20.489 ug/l  
RT: 6.342 min Scan# 80  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion: 43 Resp: 7298  
Ion Ratio Lower Upper  
43 100  
61 22.3 15.2 22.8  
87 16.3 9.8 14.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#44  
Trichloroethene  
Concen: 19.316 ug/l  
RT: 7.122 min Scan# 990  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Tgt Ion:130 Resp: 34505  
Ion Ratio Lower Upper  
130 100  
95 105.0 0.0 204.4





#45

1,2-Dichloropropane

Concen: 21.449 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01

Tgt Ion: 63 Resp: 3331

Ion Ratio Lower Upper

63 100





65 31.9 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023

Supervised By :Semsettin Yesilyurt 08/28/2023



#46

Dibromomethane

Concen: 21.150 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31




Tgt Ion: 93 Resp: 25881

Ion Ratio Lower Upper

93 100

95 82.1 66.6 100.0


174 95.6 69.4 104.0





#47  
Bromodichloromethane  
Concen: 20.684 ug/l  
RT: 7.823 min Scan# 1105  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 83 Resp: 48480  
Ion Ratio Lower Upper  
83 100  
85 66.6 49.8 74.6  
127 9.2 7.1 10.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#48  
Methyl methacrylate  
Concen: 20.633 ug/l  
RT: 7.689 min Scan# 1083  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31



Tgt Ion: 41 Resp: 35844  
Ion Ratio Lower Upper  
41 100  
69 94.4 59.2 88.8  
39 58.3 47.7 71.5





#49

1,4-Dioxane

Concen: 425.907 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01





Tgt Ion: 88 Resp: 1992

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 27.5  | 0.0   | 0.0   |
| 58  | 63.3  | 60.1  | 90.1  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#50

Toluene-d8

Concen: 52.631 ug/l

RT: 8.646 min Scan# 1240




Delta R.T. -0.000 min


Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Tgt Ion: 98 Resp: 295162

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 66.6  | 53.1  | 79.7  |





#51  
4-Methyl-2-Pentanone  
Concen: 108.507 ug/l  
RT: 8.573 min Scan# 1228  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :





VX0825WBSD01

Tgt Ion: 43 Resp: 23471  
Ion Ratio Lower Upper  
43 100  
58 43.8 30.4 45.6

Manual Integrations




APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#52  
Toluene  
Concen: 20.357 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Tgt Ion: 92 Resp: 84733  
Ion Ratio Lower Upper  
92 100  
91 169.9 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 20.073 ug/l

RT: 8.976 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01

Tgt Ion: 75 Resp: 5009

Ion Ratio Lower Upper

75 100





77 31.0 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023

Supervised By :Semsettin Yesilyurt 08/28/2023



#54

cis-1,3-Dichloropropene

Concen: 19.964 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31




Tgt Ion: 75 Resp: 53839


Ion Ratio Lower Upper

75 100

77 32.6 25.3 37.9


39 45.6 46.7 70.1#





#55  
1,1,2-Trichloroethane  
Concen: 22.109 ug/l  
RT: 9.152 min Scan# 1323  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 97      | 100   |       |       |
| 83      | 81.7  | 65.8  | 98.8  |
| 85      | 54.6  | 41.6  | 62.4  |
| 99      | 62.0  | 47.8  | 71.8  |

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#56  
Ethyl methacrylate  
Concen: 21.187 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 69      | 100   |       |       |
| 41      | 59.4  | 56.8  | 85.2  |
| 39      | 33.2  | 32.0  | 48.0  |





#57  
1,3-Dichloropropane  
Concen: 21.753 ug/l  
RT: 9.311 min Scan# 1173  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01



Tgt Ion: 76 Resp: 6215  
Ion Ratio Lower Upper  
76 100  
78 32.8 26.1 39.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#58  
2-Chloroethyl Vinyl ether  
Concen: 110.487 ug/l  
RT: 8.238 min Scan# 1173  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Tgt Ion: 63 Resp: 140763  
Ion Ratio Lower Upper  
63 100  
106 32.2 22.5 33.7





#59

2-Hexanone

Concen: 110.822 ug/l

RT: 9.427 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01

Tgt Ion: 43 Resp: 18278

Ion Ratio Lower Upper





43 100

58 60.1 26.2 78.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#60

Dibromochloromethane

Concen: 21.079 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX037262.D




Acq: 25 Aug 2023 11:31


Tgt Ion:129 Resp: 38038

Ion Ratio Lower Upper

129 100

127 76.2 38.8 116.4





#61  
1,2-Dibromoethane  
Concen: 21.513 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :





VX0825WBSD01

Tgt Ion:107 Resp: 3992  
Ion Ratio Lower Upper  
107 100  
109 94.5 75.2 112.8

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#62  
4-Bromofluorobenzene  
Concen: 53.966 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion: 95 Resp: 118938  
Ion Ratio Lower Upper  
95 100  
174 73.9 0.0 135.6  
176 71.5 0.0 131.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:117 Resp: 21818  
Ion Ratio Lower Upper  
117 100  
82 54.3 46.2 69.4  
119 33.2 25.5 38.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#64  
Tetrachloroethene  
Concen: 17.393 ug/l  
RT: 9.274 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion:164 Resp: 27685  
Ion Ratio Lower Upper  
164 100  
166 129.5 102.7 154.1  
129 98.5 80.4 120.6  
131 97.2 76.5 114.7





#65  
Chlorobenzene  
Concen: 20.638 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:112 Resp: 95172  
Ion Ratio Lower Upper  
112 100  
114 31.8 25.7 38.5

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#66  
1,1,1,2-Tetrachloroethane  
Concen: 21.162 ug/l  
RT: 10.164 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion:131 Resp: 36017  
Ion Ratio Lower Upper  
131 100  
133 93.4 47.1 141.4  
119 63.0 32.2 96.6





#67  
Ethyl Benzene  
Concen: 20.237 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 91 Resp: 164210  
Ion Ratio Lower Upper  
91 100  
106 32.0 24.9 37.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#68  
m/p-Xylenes  
Concen: 40.913 ug/l  
RT: 10.299 min Scan# 1511  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31



Tgt Ion:106 Resp: 128203  
Ion Ratio Lower Upper  
106 100  
91 206.8 165.5 248.3









#69  
o-Xylene  
Concen: 20.769 ug/l  
RT: 10.640 min Scan# 1567  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01

Tgt Ion:106 Resp: 6421  
Ion Ratio Lower Upper  
106 100  
91 219.0 108.9 326.8




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#70  
Styrene  
Concen: 21.248 ug/l  
RT: 10.658 min Scan# 1570  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31


Tgt Ion:104 Resp: 108210  
Ion Ratio Lower Upper  
104 100  
78 51.0 43.6 65.4  
103 53.9 43.6 65.4





#71  
Bromoform  
Concen: 20.573 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:173 Resp: 2619  
Ion Ratio Lower Upper  
173 100  
175 48.5 24.1 72.4  
254 0.0 0.0 0.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion:152 Resp: 113465  
Ion Ratio Lower Upper  
152 100  
115 72.8 43.3 129.9  
150 162.7 0.0 345.0





#73  
Isopropylbenzene  
Concen: 20.573 ug/l  
RT: 10.963 min Scan# 1600  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:105 Resp: 169670  
Ion Ratio Lower Upper  
105 100  
120 26.3 12.9 38.6

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#74  
N-ethyl acetate  
Concen: 20.088 ug/l  
RT: 10.841 min Scan# 1600  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 43 Resp: 64374  
Ion Ratio Lower Upper  
43 100  
70 50.3 32.9 49.3#  
55 26.5 19.3 28.9  
61 27.6 18.8 28.2





#75  
1,1,2,2-Tetrachloroethane  
Concen: 21.913 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 83 Resp: 6086  
Ion Ratio Lower Upper  
83 100  
131 10.4 5.3 15.8  
85 64.2 32.0 96.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#76  
1,2,3-Trichloropropane  
Concen: 20.394 ug/l m  
RT: 11.237 min Scan# 1665  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 75 Resp: 49281  
Ion Ratio Lower Upper  
75 100  
77 43.4 19.1 57.5





#77  
Bromobenzene  
Concen: 20.082 ug/l  
RT: 11.195 min Scan# 1658  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:156 Resp: 40448  
Ion Ratio Lower Upper  
156 100  
77 163.5 88.4 265.3  
158 98.1 48.0 144.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#78  
n-propylbenzene  
Concen: 20.993 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 91 Resp: 192791  
Ion Ratio Lower Upper  
91 100  
120 23.2 11.5 34.4





#79  
2-Chlorotoluene  
Concen: 20.547 ug/l  
RT: 11.365 min Scan# 1105  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 91 Resp: 120584  
Ion Ratio Lower Upper  
91 100  
126 33.6 16.3 48.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#80  
1,3,5-Trimethylbenzene  
Concen: 21.392 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion:105 Resp: 147168  
Ion Ratio Lower Upper  
105 100  
120 48.5 24.1 72.3





#81  
trans-1,4-Dichloro-2-butene  
Concen: 18.022 ug/l  
RT: 11.018 min Scan# 1101  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion: 75 Resp: 15592  
Ion Ratio Lower Upper  
75 100  
53 104.3 90.5 135.7  
89 48.6 36.9 55.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#82  
4-Chlorotoluene  
Concen: 20.557 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 91 Resp: 139258  
Ion Ratio Lower Upper  
91 100  
126 29.0 14.2 42.6





#83  
tert-Butylbenzene  
Concen: 22.333 ug/l  
RT: 11.713 min Scan# 1743  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:119 Resp: 146528  
Ion Ratio Lower Upper  
119 100  
91 60.1 31.2 93.6  
134 23.4 11.7 35.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#84  
1,2,4-Trimethylbenzene  
Concen: 21.188 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31



Tgt Ion:105 Resp: 147317  
Ion Ratio Lower Upper  
105 100  
120 43.6 21.2 63.6






#85  
 sec-Butylbenzene  
 Concen: 23.329 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. -0.000 min  
 Lab File: VX037262.D  
 Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01






Tgt Ion:105 Resp: 17878  
 Ion Ratio Lower Upper  
 105 100  
 134 20.1 9.8 29.4


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
 Supervised By :Semsettin Yesilyurt 08/28/2023




#86  
 p-Isopropyltoluene  
 Concen: 22.900 ug/l  
 RT: 12.006 min Scan# 1791  
 Delta R.T. -0.000 min  
 Lab File: VX037262.D  
 Acq: 25 Aug 2023 11:31



Tgt Ion:119 Resp: 151075  
 Ion Ratio Lower Upper  
 119 100  
 134 25.8 12.8 38.3  
 91 25.1 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 20.892 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01





Tgt Ion:146 Resp: 78785

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 111 | 42.3  | 22.1  | 66.1  |
| 148 | 61.4  | 31.8  | 95.3  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#88

1,4-Dichlorobenzene

Concen: 20.271 ug/l

RT: 12.042 min Scan# 1797




Delta R.T. -0.000 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Tgt Ion:146 Resp: 78975

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 111 | 43.4  | 21.3  | 63.7  |
| 148 | 64.6  | 31.8  | 95.4  |





#89

n-Butylbenzene

Concen: 23.015 ug/l

RT: 12.335 min Scan# 1844

Delta R.T. 0.006 min

Lab File: VX037262.D

Acq: 25 Aug 2023 11:31

Instrument :

MSVOA\_X

ClientSampleId :

VX0825WBSD01

Tgt Ion: 91 Resp: 125884

Ion Ratio Lower Upper

91 100

92 54.0 27.4 82.0





134 26.1 12.6 37.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023

Supervised By :Semsettin Yesilyurt 08/28/2023



#90

Hexachloroethane

Concen: 22.504 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX037262.D




Acq: 25 Aug 2023 11:31

Tgt Ion:117 Resp: 24780

Ion Ratio Lower Upper

117 100


201 71.5 32.9 98.6





#91  
1,2-Dichlorobenzene  
Concen: 21.444 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:146 Resp: 7918  
Ion Ratio Lower Upper  
146 100  
111 44.6 22.4 67.0  
148 63.5 31.6 95.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 21.785 ug/l  
RT: 12.938 min Scan# 1944  
Delta R.T. -0.006 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion: 75 Resp: 14059  
Ion Ratio Lower Upper  
75 100  
155 79.7 36.1 108.5  
157 102.1 44.9 134.7





#93  
1,2,4-Trichlorobenzene  
Concen: 22.388 ug/l  
RT: 13.585 min Scan# 2050  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:180 Resp: 4837  
Ion Ratio Lower Upper  
180 100  
182 94.0 48.3 144.8  
145 33.4 18.0 54.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023




#94  
Hexachlorobutadiene  
Concen: 24.567 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Tgt Ion:225 Resp: 20506  
Ion Ratio Lower Upper  
225 100  
223 63.4 31.1 93.2  
227 63.0 32.5 97.5





#95  
Naphthalene  
Concen: 18.216 ug/l  
RT: 13.774 min Scan# 2081  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0825WBSD01




Tgt Ion:128 Resp: 172408  
Ion Ratio Lower Upper  
128 100  
127 12.8 10.3 15.5  
129 11.5 8.8 13.2

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 08/28/2023  
Supervised By :Semsettin Yesilyurt 08/28/2023



#96  
1,2,3-Trichlorobenzene  
Concen: 22.402 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX037262.D  
Acq: 25 Aug 2023 11:31



Tgt Ion:180 Resp: 48089  
Ion Ratio Lower Upper  
180 100  
182 95.3 46.9 140.6  
145 35.1 17.8 53.5

