

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082525\
 Data File : VX047519.D
 Acq On : 25 Aug 2025 15:30
 Operator : JC/MD
 Sample : Q2954-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BRIDGE

Quant Time: Aug 26 00:39:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	200566	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	402418	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	392485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	191604	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158657	56.522	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.040%
35) Dibromofluoromethane	5.391	113	132726	50.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.640%
50) Toluene-d8	8.647	98	450174	48.224	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.440%
62) 4-Bromofluorobenzene	11.079	95	182239	52.903	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082525\
Data File : VX047519.D
Acq On : 25 Aug 2025 15:30
Operator : JC/MD
Sample : Q2954-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BRIDGE

Quant Time: Aug 26 00:39:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

