

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082620\
 Data File : VX018070.D
 Acq On : 26 Aug 2020 12:30
 Operator : JC/SP
 Sample : PB131195ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131195ZHE#01

Quant Time: Aug 27 04:33:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	454577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	718244	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	680335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	324365	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	305638	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	5.47	113	241837	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	8.70	98	848665	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
62) 4-Bromofluorobenzene	11.12	95	318773	44.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.94%	
Target Compounds						
16) Acetone	2.42	43	40090	16.303	ug/l	97
18) Methyl Acetate	2.75	43	9095	1.533	ug/l #	90
20) Methylene Chloride	2.84	84	23388	3.567	ug/l	97
43) Isopropyl Acetate	6.42	43	130079	10.467	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082620\
Data File : VX018070.D
Acq On : 26 Aug 2020 12:30
Operator : JC/SP
Sample : PB131195ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131195ZHE#01

Quant Time: Aug 27 04:33:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

