

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082620\
 Data File : VX018074.D
 Acq On : 26 Aug 2020 14:02
 Operator : JC/SP
 Sample : PB131195ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131195ZHE#05

Quant Time: Aug 27 04:38:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	480873	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	720586	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	700481	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	329306	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	336502	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
35) Dibromofluoromethane	5.46	113	267485	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
50) Toluene-d8	8.70	98	895303	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	11.12	95	360655	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
Target Compounds						
16) Acetone	2.42	43	44806	17.225	ug/l	91
18) Methyl Acetate	2.75	43	8927	1.422	ug/l #	90
20) Methylene Chloride	2.84	84	29005	4.388	ug/l	97
43) Isopropyl Acetate	6.42	43	136016	10.909	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082620\
Data File : VX018074.D
Acq On : 26 Aug 2020 14:02
Operator : JC/SP
Sample : PB131195ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131195ZHE#05

Quant Time: Aug 27 04:38:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

