

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082620\
 Data File : VX018076.D
 Acq On : 26 Aug 2020 14:47
 Operator : JC/SP
 Sample : PB131195ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131195ZHE#07

Quant Time: Aug 27 04:41:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	464870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	688535	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	667114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	301908	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	310022	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
35) Dibromofluoromethane	5.47	113	258644	55.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.32%	
50) Toluene-d8	8.70	98	863296	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
62) 4-Bromofluorobenzene	11.12	95	314242	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
Target Compounds						
16) Acetone	2.42	43	43077	17.130	ug/l	91
18) Methyl Acetate	2.75	43	9339	1.539	ug/l	95
20) Methylene Chloride	2.84	84	27078	4.196	ug/l	97
43) Isopropyl Acetate	6.42	43	130179	10.927	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082620\
Data File : VX018076.D
Acq On : 26 Aug 2020 14:47
Operator : JC/SP
Sample : PB131195ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131195ZHE#07

Quant Time: Aug 27 04:41:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

