

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082620\
 Data File : VX018083.D
 Acq On : 26 Aug 2020 17:28
 Operator : JC/SP
 Sample : L3784-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RO-06

Quant Time: Aug 27 04:51:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	467790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	739733	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	733975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	340912	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	327728	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
35) Dibromofluoromethane	5.47	113	260391	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
50) Toluene-d8	8.70	98	937725	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.12	95	334323	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
16) Acetone	2.42	43	74202	29.323	ug/l	98
18) Methyl Acetate	2.75	43	7247	1.187	ug/l #	86
20) Methylene Chloride	2.84	84	38254	6.374	ug/l	95
43) Isopropyl Acetate	6.42	43	106662	8.333	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082620\
Data File : VX018083.D
Acq On : 26 Aug 2020 17:28
Operator : JC/SP
Sample : L3784-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RO-06

Quant Time: Aug 27 04:51:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

