

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
 Data File : VX023955.D
 Acq On : 26 Aug 2021 10:20
 Operator : JC/MD
 Sample : VX0826MBL01
 Misc : 5.00G/10 mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0826MBL01

Quant Time: Aug 27 05:53:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	182187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	303718	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	273314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	134509	52.899	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.800%	
35) Dibromofluoromethane	5.397	113	98222	48.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.720%	
50) Toluene-d8	8.653	98	347677	45.442	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 90.880%#	
62) 4-Bromofluorobenzene	11.085	95	122475	43.484	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 86.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
Data File : VX023955.D
Acq On : 26 Aug 2021 10:20
Operator : JC/MD
Sample : VX0826MBL01
Misc : 5.00G/10 mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0826MBL01

Quant Time: Aug 27 05:53:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

