

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
 Data File : VX023960.D
 Acq On : 26 Aug 2021 13:11
 Operator : JC/MD
 Sample : M3525-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-FB-082221

Quant Time: Aug 27 05:55:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	249451	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102043	50.257	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.520%
35) Dibromofluoromethane	5.397	113	81218	49.192	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	8.653	98	292603	46.563	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.120%
62) 4-Bromofluorobenzene	11.085	95	105574	45.638	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.280%
Target Compounds						Qvalue
16) Acetone	2.385	43	3480	3.833	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
Data File : VX023960.D
Acq On : 26 Aug 2021 13:11
Operator : JC/MD
Sample : M3525-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-FB-082221

Quant Time: Aug 27 05:55:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

