

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
 Data File : VX023964.D
 Acq On : 26 Aug 2021 14:45
 Operator : JC/MD
 Sample : M3525-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-SW-3-GW-082221

Quant Time: Aug 27 05:56:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95244	51.001	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	102.000%	
35) Dibromofluoromethane	5.397	113	74292	48.778	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.560%	
50) Toluene-d8	8.653	98	272664	47.036	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.080%	
62) 4-Bromofluorobenzene	11.085	95	90245	42.289	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	84.580%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.892	101	59052	24.741	ug/l	98
12) 1,1-Dichloroethene	2.325	96	4684	3.648	ug/l	97
16) Acetone	2.386	43	2997	3.589	ug/l #	73
27) cis-1,2-Dichloroethene	4.495	96	3690	2.261	ug/l	91
30) Chloroform	5.105	83	43394	15.624	ug/l	99
32) 1,1,1-Trichloroethane	5.397	97	1211	0.509	ug/l #	38
44) Trichloroethene	7.135	130	76423	43.903	ug/l	98
64) Tetrachloroethene	9.275	164	2083	1.243	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
Data File : VX023964.D
Acq On : 26 Aug 2021 14:45
Operator : JC/MD
Sample : M3525-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-SW-3-GW-082221

Quant Time: Aug 27 05:56:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

