

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
 Data File : VX023970.D
 Acq On : 26 Aug 2021 17:07
 Operator : JC/MD
 Sample : M3535-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-SW-05-GW-082321

Quant Time: Aug 27 05:58:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234013	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	101711	53.935	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.880%			
35) Dibromofluoromethane	5.397	113	78330	50.572	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.140%			
50) Toluene-d8	8.653	98	285822	48.485	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 96.960%			
62) 4-Bromofluorobenzene	11.085	95	104069	47.955	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.900%			
Target Compounds						Qvalue
7) Trichlorofluoromethane	1.892	101	7500	3.112	ug/l	96
12) 1,1-Dichloroethene	2.319	96	677	0.522	ug/l	94
16) Acetone	2.386	43	2461	2.919	ug/l #	80
30) Chloroform	5.111	83	6564	2.340	ug/l	92
44) Trichloroethene	7.135	130	10269	5.801	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
Data File : VX023970.D
Acq On : 26 Aug 2021 17:07
Operator : JC/MD
Sample : M3535-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-SW-05-GW-082321

Quant Time: Aug 27 05:58:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

