

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
 Data File : VX023971.D
 Acq On : 26 Aug 2021 17:30
 Operator : JC/MD
 Sample : M3535-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WVLY-SW-14-GW-082321

Quant Time: Aug 27 05:58:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	252572	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106129	51.118	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.240%
35) Dibromofluoromethane	5.397	113	82860	49.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.140%
50) Toluene-d8	8.653	98	304164	47.805	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.600%
62) 4-Bromofluorobenzene	11.085	95	109050	46.558	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.120%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2727	2.938	ug/l	93
30) Chloroform	5.105	83	3783	1.225	ug/l #	78
44) Trichloroethene	7.129	130	820	0.429	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082621\
Data File : VX023971.D
Acq On : 26 Aug 2021 17:30
Operator : JC/MD
Sample : M3535-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WVLY-SW-14-GW-082321

Quant Time: Aug 27 05:58:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

