

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082625\
 Data File : VX047531.D
 Acq On : 26 Aug 2025 13:30
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 27 01:36:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	225174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	453719	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	457530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	229596	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	172613	54.774	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.540%
35) Dibromofluoromethane	5.397	113	143266	48.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.300%
50) Toluene-d8	8.653	98	515696	48.997	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.079	95	217855	56.091	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082625\
Data File : VX047531.D
Acq On : 26 Aug 2025 13:30
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 27 01:36:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

