

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082718\
 Data File : VX004261.D
 Acq On : 27 Aug 2018 13:47
 Operator : JC/MD
 Sample : VX0827MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0827MBL01

Quant Time: Aug 27 16:33:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 14:20:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	288367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	450295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	425678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	230882	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	192828	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
35) Dibromofluoromethane	5.50	113	170746	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
50) Toluene-d8	8.71	98	628149	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
62) 4-Bromofluorobenzene	11.14	95	219109	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082718\
Data File : VX004261.D
Acq On : 27 Aug 2018 13:47
Operator : JC/MD
Sample : VX0827MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0827MBL01

Quant Time: Aug 27 16:33:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 14:20:03 2018
Response via : Initial Calibration

