

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082720\
 Data File : VX018095.D
 Acq On : 27 Aug 2020 16:41
 Operator : JC/SP
 Sample : PB131126TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131126TB

Quant Time: Aug 28 03:55:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	558482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	893274	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	867905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	417775	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	381040	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
35) Dibromofluoromethane	5.47	113	297428	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	8.70	98	1114637	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
62) 4-Bromofluorobenzene	11.12	95	418598	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
Target Compounds						
16) Acetone	2.42	43	48944	16.201	ug/l	92
18) Methyl Acetate	2.75	43	10537	1.445	ug/l	92
20) Methylene Chloride	2.83	84	15455	1.366	ug/l	91
43) Isopropyl Acetate	6.42	43	162313	10.502	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082720\
Data File : VX018095.D
Acq On : 27 Aug 2020 16:41
Operator : JC/SP
Sample : PB131126TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131126TB

Quant Time: Aug 28 03:55:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

