

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082720\
 Data File : VX018100.D
 Acq On : 27 Aug 2020 18:36
 Operator : JC/SP
 Sample : L3820-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB181-GW-638-640

Quant Time: Aug 28 09:31:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	594052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	955502	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	917106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	426035	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	398297	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
35) Dibromofluoromethane	5.47	113	314413	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
50) Toluene-d8	8.70	98	1197413	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.12	95	444076	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
Target Compounds						
16) Acetone	2.42	43	22163	6.897	ug/l	92
17) Carbon Disulfide	2.55	76	7935	0.452	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082720\
Data File : VX018100.D
Acq On : 27 Aug 2020 18:36
Operator : JC/SP
Sample : L3820-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB181-GW-638-640

Quant Time: Aug 28 09:31:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

